

# Development of a GRUAN data product for Cryogenic Frostpoint Hygrometers

Holger Vömel,  
GRUAN Lead Center  
DWD Meteorological Observatory  
Lindenberg

ICM-6  
10 March 2014



## From manufacturer:

- Measurement range: **-100°C to +25°C**
- Range for which the sensor was calibrated: **-95°C to +18°C**
- Calibration accuracy : **<0.05 K throughout measurement range**
- Time resolution of the data transmission: **1 s**
- Accuracy, precision, repeatability, or uncertainty: **Systematic errors < 0.5 K**
- Time lag constants: **Not well understood**
- Is a recalibration required as part of the launch preparation? **No**

## From site:

- Tests of a limited number of sondes in an environmental chamber using reference sensors:

**AquaVIT 1 and 2 (Stratospheric uncertainty < 10% mr)**

- Tests of a larger number of sondes at relaxed conditions

**Some lab tests in Lindenberg with about 15 sondes.**

- Test of the production variability

**Some lab tests in Lindenberg with about 15 sondes.**

- Dual launches with a GRUAN accepted sonde

**Every sonde launched with Vaisala RS92**

- Manufacturer independent ground check

**In progress**

## Random uncertainty:


- + Related to statistical variations in the entire measurement system
- + Can be measured by the scatter of multiple measurements under identical conditions
- + Averages out in the mean of a large number of observations.
- + Often called precision

## Systematic error:


- + Deviation from the (unknown) true value due to (unspecified) instrument artifacts
- + Can only be estimated
- + Does not average out in the mean of a large number of observations

Thermistor calibration :	<0.05 K	(<1% mixing ratio)
systematic in profile		
systematic in time series		
Manufacturing variability :	< 0.2 K	(< 4% mixing ratio)
systematic in profile		
random in time series		
Controller stability :	0.1 K ... 0.5 K	(2% to 10% mixing ratio)
random in profile		(occasionally larger)
random in time series		
PID drift : 0 K		
systematic in profile		
random in time series		
Contamination :	Can be large. Is flagged out in processing	


# CFH Repeatability



# PID Stability





# Controller stability







# Uncertainty and resolution



# Uncertainty and resolution



# Next steps

- Implement ground check
- Complete processing routines for CFH
- Merge with RS92 data product
- Publication